



PerkinElmer, Inc.  
549 Albany Street  
Boston, MA 02118 USA

P: (800) 762-4000  
(+1) 203-925-4602  
www.perkinelmer.com

## CERTIFICATE OF ANALYSIS

### NeoMDx™ cCMV real-time PCR assay (CMV-96 and CMV-384)

#### NeoMDx™ cCMV PCR Reagent Kit

Catalog No: CMV-RGT-96 and CMV-RGT-384  
Lot No: 3012096  
Kit Expiration Date: 20-Jan-2024


Component	Lot No.
cCMV PCR Reagent 1	3012101
cCMV PCR Reagent 2	3012102

Items	Criteria	Results	Disposition
CMV Sensitivity Panel	Sensitivity panel must be detected as 3/3	Standards compliant	<input checked="" type="checkbox"/> Pass <input type="checkbox"/> Fail
RPP30 Sensitivity Panel	Sensitivity panel must be detected as 3/3	Standards compliant	<input checked="" type="checkbox"/> Pass <input type="checkbox"/> Fail
CMV Standard Curve	Standard Curve Slope and y-intercept fit QC	Standards compliant	<input checked="" type="checkbox"/> Pass <input type="checkbox"/> Fail
RPP30 Standard Curve	Standard Curve Slope and y-intercept fit QC	Standards compliant	<input checked="" type="checkbox"/> Pass <input type="checkbox"/> Fail
Negative Control Panel	All targets are negative based on Ct cutoff	Standards compliant	<input checked="" type="checkbox"/> Pass <input type="checkbox"/> Fail

Store -15°C to -20°C.

For research use only. Not for use in diagnostic procedures.

The above listed component lot of NeoMDx™ cCMV real-time PCR assay (CMV-96 and CMV-384) has been tested and approved by Quality.

Approved by:  \_\_\_\_\_  
CD79562994D14DB...

20 April 2022 | 2:06 PM EDT

Date: \_\_\_\_\_

For a complete listing of our global offices, visit [www.perkinelmer.com/ContactUs](http://www.perkinelmer.com/ContactUs)